

**Components:**

- #130-31 Thermostat
- #130-38-2 Red Lens for Lamp
- #130-38-3 Lamp
- #130-38-7 Stainless Steel Cup
- #171-32 Knob

*Optional:*

An additional power cord is necessary when using the Universal Heat Cup without a Model 900 Viscometer.

- #152-37 AC Power Cord, 115-Volt
- #152-38 AC Power Cord, 230-Volt

**OFI Testing Equipment, Inc.**

11302 Steeplecrest Dr. Houston, TX 77065 USA  
 Tele: 713.880.9885 or 877.837.8683  
 Fax: 713.880.9886 www.ofite.com

©Copyright OFITE 2011



*Dependable Products From People You Trust*



**Universal Heat Cup**  
**No. 130-76-10 - 115 Volt**  
**No. 130-76-10-1 - 230 Volt**

**Instruction Manual**

Updated 5/25/2011  
Ver. 1.0

**OFI Testing Equipment, Inc.**

**Components:**

- #130-31 Thermostat
- #130-38-2 Red Lens for Lamp
- #130-38-3 Lamp
- #130-38-7 Stainless Steel Cup
- #171-32 Knob

*Optional:*

An additional power cord is necessary when using the Universal Heat Cup without a Model 900 Viscometer.

- #152-37 AC Power Cord, 115-Volt
- #152-38 AC Power Cord, 230-Volt

**OFI Testing Equipment, Inc.**

11302 Steeplecrest Dr. Houston, TX 77065 USA  
 Tele: 713.880.9885 or 877.837.8683  
 Fax: 713.880.9886 www.ofite.com

©Copyright OFITE 2011



*Dependable Products From People You Trust*



**Universal Heat Cup**  
**No. 130-76-10 - 115 Volt**  
**No. 130-76-10-1 - 230 Volt**

**Instruction Manual**

Updated 5/25/2011  
Ver. 1.0

**OFI Testing Equipment, Inc.**

**Introduction:**

The OFITE Universal Heat Cup is designed for controlling the temperature of a fluid sample while taking readings with a rheometer or viscometer. Normal heat-up time is 15 minutes and the pilot light turns on during heating. Drilling fluid has a low thermal conductivity, so it must be agitated in order to reach a uniform temperature within a reasonable length of time.

The Universal Heat Cup is designed to work with the OFITE Model 900 Viscometer. It can be plugged directly into the Viscometer, eliminating the need for a second power outlet. This also allows the ORCAD software to control the temperature of the fluid sample during a test.

It can also be used with other viscometers like the OFITE Model 800 Viscometer and the OFITE Hand-Crank Rheometer. An additional power cable is necessary (not included).

Refer to the viscometer instruction manual for details on using the viscometer.

**Introduction:**

The OFITE Universal Heat Cup is designed for controlling the temperature of a fluid sample while taking readings with a rheometer or viscometer. Normal heat-up time is 15 minutes and the pilot light turns on during heating. Drilling fluid has a low thermal conductivity, so it must be agitated in order to reach a uniform temperature within a reasonable length of time.

The Universal Heat Cup is designed to work with the OFITE Model 900 Viscometer. It can be plugged directly into the Viscometer, eliminating the need for a second power outlet. This also allows the ORCAD software to control the temperature of the fluid sample during a test.

It can also be used with other viscometers like the OFITE Model 800 Viscometer and the OFITE Hand-Crank Rheometer. An additional power cable is necessary (not included).

Refer to the viscometer instruction manual for details on using the viscometer.

**Operation:**

1. Plug the power cord into a suitable power outlet (either 115 or 230 volt) or into the back of a Model 900 Viscometer.

When plugging directly into a power outlet, an additional power cord is necessary (#152-37 for 115 volt or #152-38 for 230 volt).

2. Add test fluid to the stainless steel cup and place it in the heating well.
3. Immerse the rotor in test fluid to the scribed line. Insert a thermometer or thermocouple into the fluid.
4. Place the viscometer on the stir setting.
5. Set the thermostat about 2/3 of the way to the highest setting. Wait for the temperature to stabilize. Adjust the thermostat if necessary.

If you are using the ORCAD software, turn the thermostat on the Heat Cup up to the highest setting. The software will control the temperature.

**Operation:**

1. Plug the power cord into a suitable power outlet (either 115 or 230 volt) or into the back of a Model 900 Viscometer.

When plugging directly into a power outlet, an additional power cord is necessary (#152-37 for 115 volt or #152-38 for 230 volt).

2. Add test fluid to the stainless steel cup and place it in the heating well.
3. Immerse the rotor in test fluid to the scribed line. Insert a thermometer or thermocouple into the fluid.
4. Place the viscometer on the stir setting.
5. Set the thermostat about 2/3 of the way to the highest setting. Wait for the temperature to stabilize. Adjust the thermostat if necessary.

If you are using the ORCAD software, turn the thermostat on the Heat Cup up to the highest setting. The software will control the temperature.

**Maintenance:**

Thoroughly wash the removable stainless steel cup with soap and water after each use.

**Caution:**

1. Always unplug the Universal Heat Cup (or turn off the Model 900 Viscometer) when not in use.
2. Never heat fluid over 200°F (93° C).
3. Never immerse the Universal Heat Cup in water when cleaning.

**Maintenance:**

Thoroughly wash the removable stainless steel cup with soap and water after each use.

**Caution:**

1. Always unplug the Universal Heat Cup (or turn off the Model 900 Viscometer) when not in use.
2. Never heat fluid over 200°F (93° C).
3. Never immerse the Universal Heat Cup in water when cleaning.