



294-50 Capillary Suction Timer Calibration Report

Order Number: 167775
 Serial Number: 529220000002

Calibration Date: 9/20/2024
 Calibration Due Date: 9/20/2025

Fluid Information

Fluid μ S	Fluid Lot Number	Exp. Date
10000 μ S	4061201	6/30/2025

Total dissolved of solids = 6600 Parts Per Million (PPM)

Thermometer I.D. # 260
 Model Number: 90080-09
 Serial Number: 192283369

Calibration Due Date: 1/29/2025

Calibration Check

Calibration Check		Temp (°F)
Reading 1 (Sec)	13.6 Sec	73.7°F
Reading 2 (Sec)	12.5 Sec	73.7°F
Reading 3 (Sec)	11.7 Sec	73.7°F
Max Deviation Allowed	\pm 3.6 Sec	
Actual Deviation (Sec)	1.9 Sec	
Pass / Fail	Pass	

Three tests of the same temperature should produce a deviation of no more than \pm 3.6 second:

Technician Tested: Magdy Beshay

Date: 9/20/2024

Approved By: Samantha Korenek

Date: 9/23/2024