



112-00-1-99 Reconditioned 230V  
Calibration Report

Order Number:	536868	Calibration Date:	9/3/2024
Serial Number:	536868000001	Recommended Calibration Date:	9/3/2025
Part Number:	112-00-1-99	Manufacturer:	OFITE
Item I.D. #:	NA		

RPM Calibration Check:				
Target Speed	Tachometer Reading (RPM)	Allowed Deviation	Actual Deviation	Pass / Fail
60 RPM	59.7	± 1.5 RPM	.3 RPM	Pass
200 RPM	199.7	± 5.0 RPM	.3 RPM	Pass
600 RPM	599.1	± 15.0 RPM	.9 RPM	Pass
1000 RPM	998.9	± 25.0 RPM	1.1 RPM	Pass

**Tachometer Calibration Information:** 708

Model Number: <span style="border-bottom: 1px solid black;">R7050</span>	Calibration Due Date: <span style="border-bottom: 1px solid black;">1/30/2025</span>
Serial Number: <span style="border-bottom: 1px solid black;">230832873</span>	

Deionized Water Check				
Target Torque (in-lb)	Unit Reading (in-lb)	Acceptable Deviation (in-lb)	Actual Deviation	Pass / Fail
34.0 in-lb	33.9	± 2.0 in -lb	0.1 in-lb	Pass

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: <span style="border-bottom: 1px solid black;">Magdy Beshay</span>	Date: <span style="border-bottom: 1px solid black;">9/3/2024</span>
Approved By: <span style="border-bottom: 1px solid black;">Samantha Korenek</span>	Date: <span style="border-bottom: 1px solid black;">9/5/2024</span>

