



EP and Lubricity Tester Calibration Report

Order Number: 22133
 Serial Number: 09-01
 Part Number: 111-00-99
 Item I.D. #: NA

Calibration Date: 10/1/2024
 Recommended Calibration Date: 10/1/2025
 Manufacturer: OFI

| RPM Calibration Check | | | | |
|-----------------------|--------------------------|-------------------|------------------|-------------|
| Target Speed | Tachometer Reading (RPM) | Allowed Deviation | Actual Deviation | Pass / Fail |
| 60 RPM | 59.0 | ± 1.5 RPM | 1.0 RPM | Pass |
| 200 RPM | 200.9 | ± 5.0 RPM | .9 RPM | Pass |
| 600 RPM | 601.0 | ± 15.0 RPM | 1.0 RPM | Pass |
| 1000 RPM | 1002.0 | ± 25.0 RPM | 2.0 RPM | Pass |

Tachometer Calibration Information: 155

Model Number: 4060 Calibration Due Date: 7/22/2025
 Serial Number: 122574032

| Deionized Water Check | | | | |
|-----------------------|----------------------|------------------------------|------------------|-------------|
| Target Torque (in-lb) | Unit Reading (in-lb) | Acceptable Deviation (in-lb) | Actual Deviation | Pass / Fail |
| 34.0 in-lb | 32.5 | ± 2.0 in -lb | 1.5 in-lb | Pass |

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Emad Hanna Date: 10/1/2024

Approved By: Samantha Korenek Date: 10/1/2024