



## 100-70 Pressurized Mud Balance Calibration Report

Order Number: 22190      Calibration Date: 10/3/2024

Serial Number: 1130810      Recommended Due Date: 10/3/2025

Density Calibration			
Target	Reading	Allowed Deviation	Pass / Fail
1.0 g/cm <sup>3</sup>	1.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	Pass
2.0 g/cm <sup>3</sup>	2.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	Pass

<b>Ambient Monitor</b>	345
Model Number:	RH300
Serial Number:	Z395285
Calibration Due:	8/20/2025

<b>Ambient Conditions</b>	
Lab Temperature:	73.9°F
Lab Humidity:	45.5%

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master monitoring equipment that are traceable to NIST.

Technician Tested: Elmer Pumphrey

Date: 10/3/2024

Approved By: Samantha Korenek

Date: 10/4/2024