



## EP and Lubricity Tester Calibration Report

Order Number: 168801  
 Serial Number: 13-62  
 Part Number: 112-00  
 Item I.D. #: NA

Calibration Date: 11/18/2024  
 Recommended Calibration Date: 11/18/2025  
 Manufacturer: OFITE

RPM Calibration Check				
Target Speed	Tachometer Reading (RPM)	Allowed Deviation	Actual Deviation	Pass / Fail
60 RPM	59.6	± 1.5 RPM	.4 RPM	Pass
200 RPM	199.8	± 5.0 RPM	.2 RPM	Pass
600 RPM	600.1	± 15.0 RPM	.1 RPM	Pass
1000 RPM	1000.3	± 25.0 RPM	.3 RPM	Pass

**Tachometer Calibration Information:** 155

Model Number: 4060 Calibration Due Date: 7/22/2025  
 Serial Number: 122574032

Deionized Water Check				
Target Torque (in-lb)	Unit Reading (in-lb)	Acceptable Deviation (in-lb)	Actual Deviation	Pass / Fail
34.0 in-lb	34.3	± 2.0 in -lb	0.3 in-lb	Pass

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Emad Hanna Date: 11/18/2024

Approved By: Samantha Korenek Date: 11/18/2024