



Dependable Products From People You Trust

# OFI TESTING EQUIPMENT, INC.

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com


## CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

AS FOUND CHART

AS LEFT CHART

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
65.00 °F	60.00 %	165-14-1	5374970000004	930.0 °F	972.1 °F	860-1000 °F	12/18/2024	12/18/2025	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

Channel 4

**CAL STATION SERIAL NUMBER:** 6B-11    **CAL STATION CALIBRATION DATE:** 11/11/2024    **CAL STATION DUE DATE:** 11/11/2025

**AMBIENT MONITOR MODEL:** RH101    **AMBIENT MONITOR S/N:** Z375174    **AMBIENT MONITOR CAL DATE:** 01/16/2024    **AMBIENT MONITOR DUE DATE:** 01/16/2025

**UNCERTAINTY:** ± 2.8°C / 5°F    **CALIBRATION PROCEDURE USED** Retort Calibration Procedures API RP 13B-1

**AS FOUND CONDITION:**

**CALIBRATION RESULT:** Accepted