



Dependable Products From People You Trust

OFI TESTING EQUIPMENT, INC.


11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com

CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

- AS FOUND CHART
- AS LEFT CHART

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
65.00 °F	60.00 %	165-14-1	5374970000008	930.0 °F	972.3 °F	860-1000 °F	12/18/2024	12/18/2025	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

Channel 8

CAL STATION SERIAL NUMBER: 6B-11
CAL STATION CALIBRATION DATE: 11/11/2024
CAL STATION DUE DATE: 11/11/2025
AMBIENT MONITOR MODEL: RH101
AMBIENT MONITOR S/N: Z375174
AMBIENT MONITOR CAL DATE: 01/16/2024
AMBIENT MONITOR DUE DATE: 01/16/2025
UNCERTAINTY: ± 2.8°C / 5°F
CALIBRATION PROCEDURE USED: Retort Calibration Procedures API RP 13B-1

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted