



Dependable Products From People You Trust

OFI TESTING EQUIPMENT, INC.


11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com

CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

- AS FOUND CHART
- AS LEFT CHART

LAB TEMPERATURE 65.00 °F	LAB HUMIDITY 60.00 %	PART NUMBER 165-14-1	SERIAL NUMBER 537844000003	SET POINT 930.0 °F	RESULT 993.7 °F	ALLOWABLE RANGE 860-1000 °F	CALIBRATION DATE 12/18/2024	DUE DATE 12/19/2025	TECHNICIAN SM
------------------------------------	--------------------------------	--------------------------------	--------------------------------------	------------------------------	---------------------------	---------------------------------------	---------------------------------------	-------------------------------	-------------------------



Calibration station has been calibrated using instruments which are traceable to NIST.

Channel 3

CAL STATION SERIAL NUMBER: 6B-11 **CAL STATION CALIBRATION DATE:** 11/11/2024 **CAL STATION DUE DATE:** 11/11/2025

AMBIENT MONITOR MODEL: RH101 **AMBIENT MONITOR S/N:** Z375174 **AMBIENT MONITOR CAL DATE:** 01/16/2024 **AMBIENT MONITOR DUE DATE:** 01/16/2025

UNCERTAINTY: ± 2.8°C / 5°F **CALIBRATION PROCEDURE USED:** Retort Calibration Procedures API RP 13B-1

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted