



**OFI Testing Equipment, Inc.**  
 11302 Steeplecrest Drive Houston, TX 77065 USA  
 Phone: 832-320-7300 · Fax: 713-880-9886 · Web: www.ofite.com

**Viscometer Calibration Report**

Oil Calibration Per API Recommended Practices; 13B-1 Annex I.5, 13B-2 Annex J.5, 10B-2 Annex B.3.6 and Specification 13A Section 5.2.6

Part Number: 130-76-C Model: Model 900 Viscometer, Complete, 115 Volt Serial Number: 5374500000004 S/O# or J/O# New  
 Bob B1 Spring F1

| RPM Check            | ± 1 RPM for speeds <100 RPM |            | ± 1 % of nominal speed of 100 RPM or Greater |           |           |           |
|----------------------|-----------------------------|------------|----------------------------------------------|-----------|-----------|-----------|
| Speed Setting (RPM): | 3 RPM                       | 6 RPM      | 100 RPM                                      | 200 RPM   | 300 RPM   | 600 RPM   |
| Actual:              | 3.0 RPM                     | 6.0 RPM    | 100.0 RPM                                    | 200.0 RPM | 300.0 RPM | 600.0 RPM |
| Allowed Deviation:   | ± 1.00 RPM                  | ± 1.00 RPM | ± 1.0 RPM                                    | ± 2.0 RPM | ± 3.0 RPM | ± 6.0 RPM |
| Actual Deviation:    | 0.0 RPM                     | 0.0 RPM    | 0.0 RPM                                      | 0.0 RPM   | 0.0 RPM   | 0.0 RPM   |
| Pass / Fail:         | Pass                        | Pass       | Pass                                         | Pass      | Pass      | Pass      |

Per API RP 10B-2 Annex B, Equipment Calibration Requirements

**AS LEFT MEASUREMENTS**

| Calibration Fluid Viscosity | Calibration Fluid Batch # | Temp | Calibrated Viscosity (CV) | 600 RPM Reading | 600 VIS: 600 Reading * 0.5050 | Accepted Deviation | Actual Deviation** | 300 RPM Reading | 300 VIS: 300 Reading * 1.0000 | Accepted Deviation | Actual Deviation | 200 RPM Reading | 200 VIS: 200 Reading / 0.6667 | Accepted Deviation | Actual Deviation | 100 RPM Reading | 100 VIS: 100 Reading / 0.3333 | Accepted Deviation | Actual Deviation (100 VIS - CV) | Pass/Fail |
|-----------------------------|---------------------------|------|---------------------------|-----------------|-------------------------------|--------------------|--------------------|-----------------|-------------------------------|--------------------|------------------|-----------------|-------------------------------|--------------------|------------------|-----------------|-------------------------------|--------------------|---------------------------------|-----------|
| 100 cP                      | 309602                    | 22.9 | 101.9                     | 203.4           | 102.72                        | ±1.50              | 0.82               | 101.9           | 101.9                         | ±1.50              | 0.00             | 68.4            | 102.59                        | ±1.50              | 0.69             | 34.4            | 103.21                        | ±1.50              | 1.31                            | Pass      |
| 50 cP                       | 228501                    | 22.5 | 50.4                      | 99.4            | 50.20                         | ±1.50              | -0.20              | 50.7            | 50.7                          | ±1.50              | 0.30             | 34.2            | 51.30                         | ±1.50              | 0.90             | 17              | 51.01                         | ±1.50              | 0.61                            | Pass      |

\*\* For viscometers not using a B1F1 configuration, the 600 rpm reading is shown, but is not used for determining a pass/fail.

|                    |      |
|--------------------|------|
| Calibration Result | Pass |
|--------------------|------|

Technician: Emad Hanna Factory Recommendations\*: Factory recalibration is recommended annually. Uncertainty: 0.5% Reviewed By: Therese Ikladious  
 Calibration Date: 2/3/2025 Calibration requirements are determined per customer's Quality Management System. Review Date: 2/3/2025  
 Calibration Due Date\*: 2/3/2026 Per API RP 13B-1 and -2, viscometers used for testing drilling fluids should be verified with a traceable calibration fluid at least monthly.  
 Per API RP 10B-2, viscometers used for testing cements should be verified for accuracy with a traceable calibration fluid at least quarterly.

Calibration Fluid Activation Date: 100 cP 2/3/2025 50 cP 2/3/2025  
 (Activation Date should not exceed 30 days of the calibration date)

|                                                                                                                                                                                      |  |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--|
| For internal use only.                                                                                                                                                               |  |
| All master instruments and fluids are calibrated and/or certified to standards which are traceable to NIST. <input type="checkbox"/> Instrument Certificates of Calibration attached |  |
| Fluid Thermometer: <u>704</u> Model #: <u>4000,90080-05</u> Calibration Date: <u>5/8/2024</u> Calibration Due: <u>5/8/2025</u> Serial No: <u>221540132</u>                           |  |
| Tachometer: <u>155</u> Model #: <u>4060</u> Calibration Date: <u>7/22/2024</u> Calibration Due: <u>7/22/2025</u> Serial No: <u>122574032</u>                                         |  |
| Ambient Monitor: <u>345</u> Model #: <u>RH300</u> Calibration Date: <u>8/20/2024</u> Calibration Due: <u>8/20/2025</u> Serial No: <u>2395285</u>                                     |  |
| Ambient Conditions                                                                                                                                                                   |  |
| Room Temperature: <u>72</u> Relative Humidity: <u>48</u>                                                                                                                             |  |