



Dependable Products From People You Trust

OFI TESTING EQUIPMENT, INC.

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com


CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

AS FOUND CHART

AS LEFT CHART

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
63.00 °F	56.00 %	165-10	5376690000003	930.0 °F	997.0 °F	860-1000 °F	12/27/2024	12/27/2026	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

Channel 1

CAL STATION SERIAL NUMBER: 6B-11 **CAL STATION CALIBRATION DATE:** 11/11/2024 **CAL STATION DUE DATE:** 11/11/2025

AMBIENT MONITOR MODEL: RH101 **AMBIENT MONITOR S/N:** Z375174 **AMBIENT MONITOR CAL DATE:** 01/16/2024 **AMBIENT MONITOR DUE DATE:** 01/16/2025

UNCERTAINTY: ± 2.8°C / 5°F **CALIBRATION PROCEDURE USED** Retort Calibration Procedures API RP 13B-1

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted