OFITE
OFI Testing Equipment, Inc.

Order Number:	RP00000132
Serial # / Local I.D. #:	16-284
Calibration Date:	8/18/2026
Recommended Due Date:	8/18/2027
Certificate Number:	16-28446252
Condition Found:	Out of Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

100-70 Pressurized Mud Balance
Calibration Report

Density Calibration Check

Target	Reading	Allowed Deviation	Actual Deviation	Pass / Fail
1.00 g/cm3	1.00 g/cm3 Pass	± 0.01 g/cm3	0.00 g/cm3	Pass
2.00 g/cm3	2.00 g/cm3 Pass	± 0.01 g/cm3	0.00 g/cm3	Pass

Ambient Monitor I.D. #	261	Certificate Number	C-59375
Model Number	EA20	Calibration Date	1/20/2026
Serial Number	Z375174	Calibration Due Date	1/20/2027

Ambient Conditions

Lab Temperature	71.2 F°	Lab Humidity	48.9 %

There is no specific pass or fail criteria per API RP 10B-2. This table is to provide a record of the current ambient conditions during the calibration process.

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Magdy Beshay Calibration Technician	Magdy Beshay Technician Signature	8/18/2026 Date
Samantha Korenek Reviewed By	Samantha Korenek Reviewer Signature	8/18/2026 Date