



100-70 Pressurized Mud Balance Calibration Report

Order Number: 22362 Calibration Date: 12/20/2024

Serial Number: 16-523 Recommended Due Date: 12/20/2025

Density Calibration			
Target	Reading	Allowed Deviation	Pass / Fail
1.0 g/cm ³	1.00 g/cm ³	± 0.01 g/cm ³	Pass
2.0 g/cm ³	2.00 g/cm ³	± 0.01 g/cm ³	Pass

Ambient Monitor	261
Model Number:	EA20
Serial Number:	Z375174
Calibration Due:	1/16/2025

Ambient Conditions	
Lab Temperature:	72.3°F
Lab Humidity:	46.1%

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master monitoring equipment that are traceable to NIST.

Technician Tested: Magdy Beshay

Date: 12/20/2024

Approved By: Samantha Korenek

Date: 12/27/2024