



## 171-49 Temperature Controller Calibration Report

Order Number: NEW Calibration Date: 1/17/2025  
 Serial Number: 5380990000001 Recommended Calibration Date: 1/17/2026

Thermocouple Calibration						
Channel	Target Temp.	As Found Temp	Deviation Allowed	Adjusted Temp	Deviation Allowed	Pass / Fail
Channel 1	120°F	116°F	± 4°F	120°F	± 0°F	Pass
	450°F	444°F	± 10°F	450°F	± 0°F	Pass

**Dry Block Calibration Information:** 061

Model Number: CL1000A Calibration Due Date: 6/12/2025  
 Serial Number: 13129620

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Shawn Maglitto Date: 01/17/2025

Approved By: Samantha Korenek Date: 01/17/2025