



## Silverson Mixer Calibration Report

Order Number: 22392  
 Part Number: 152-15  
 Serial Number: 35761  
 I.D. Number: \_\_\_\_\_

Verification Date: 1/16/2025  
 Verification Due Date: 1/16/2026

### Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2516 RPM	± 300 RPM	16 RPM	Pass
5000 RPM	5000 RPM	5022 RPM	± 300 RPM	22 RPM	Pass
7500 RPM	7500 RPM	7516 RPM	± 300 RPM	16 RPM	Pass
9200 RPM	9200 RPM	9337 RPM	± 300 RPM	137 RPM	Pass

Calibrated Tachometer I.D. 678

Model Number: R7050  
 Serial Number: 220326760

Calibration Due Date: 7/22/2025

Calibrated By: Magdy beshay

Date: 1/16/2025

Approved By: Samantha Korenek

Date: 1/20/2025

