



Silverson Mixer Calibration Report

Order Number: 22392
 Part Number: 152-15
 Serial Number: 33075
 I.D. Number: _____

Verification Date: 1/16/2025
 Verification Due Date: 1/16/2026

Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2511 RPM	± 300 RPM	11 RPM	Pass
5000 RPM	5000 RPM	5030 RPM	± 300 RPM	30 RPM	Pass
7500 RPM	7500 RPM	7531 RPM	± 300 RPM	31 RPM	Pass
9200 RPM	9200 RPM	9268 RPM	± 300 RPM	68 RPM	Pass

Calibrated Tachometer I.D. 678

Model Number: R7050
 Serial Number: 220326760

Calibration Due Date: 7/22/2025

Calibrated By: Magdy beshay

Date: 1/16/2025

Approved By: Samantha Korenek

Date: 1/20/2025

