



171-49-2 Temperature Controller Calibration Report

Order Number: 22441
 Serial Number: 8143

Calibration Date: 2/17/2025
 Recommended Calibration Date: 2/17/2026

Thermocouple Calibration						
Channel	Target Temp.	As Found Temp	Deviation Allowed	Adjusted Temp	Deviation Allowed	Pass / Fail
Channel 1	120°F	119°F	± 4°F	120°F	± 0°F	Pass
	450°F	448°F	± 10°F	450°F	± 0°F	Pass
Channel 2	120°F	119°F	± 4°F	120°F	± 0°F	Pass
	450°F	446°F	± 10°F	450°F	± 0°F	Pass

Dry Block Calibration Information: 700

Model Number: ETC-400A
 Serial Number: 704749-02572

Calibration Due Date: 4/16/2025

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay

Date: 2/17/2025

Approved By: Samantha Korenek

Date: 2/19/25