



Silverson Mixer Calibration Report

Order Number: 22488
 Part Number: 152-15
 Serial Number: 36389
 I.D. Number: _____

Verification Date: 3/4/2025
 Verification Due Date: 3/4/2026

Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2506 RPM	± 300 RPM	06 RPM	Pass
5000 RPM	5000 RPM	5027 RPM	± 300 RPM	27 RPM	Pass
7500 RPM	7500 RPM	7534 RPM	± 300 RPM	34 RPM	Pass
9200 RPM	9200 RPM	9256 RPM	± 300 RPM	56 RPM	Pass

Calibrated Tachometer I.D. 708

Model Number: R7050
 Serial Number: 230832873

Calibration Due Date: 2/6/2026

Calibrated By: Magdy Beshay

Date: 3/4/2025

Approved By: Samantha Korenek

Date: 3/11/2025

