



Dependable Products From People You Trust

OFI TESTING EQUIPMENT, INC.

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com


CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

AS FOUND CHART

AS LEFT CHART

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
74.00 °F	45.00 %	165-14-3	5384890000001	930.0 °F	894.8 °F	860-1000 °F	3/12/2025	3/12/2026	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

CAL STATION SERIAL NUMBER: 6B-11

CAL STATION CALIBRATION DATE: 11/11/2024

CAL STATION DUE DATE: 11/11/2025

AMBIENT MONITOR MODEL: RH300

AMBIENT MONITOR S/N: Z395285

AMBIENT MONITOR CAL DATE: 08/20/2024

AMBIENT MONITOR DUE DATE: 08/20/2025

UNCERTAINTY: ± 2.8°C / 5°F

CALIBRATION PROCEDURE USED: Retort Calibration Procedures API RP 13B-1

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted

- Channel 2
- Plot 1
- Plot 2
- Plot 3
- Plot 4