



Chiller Calibration Report

Order Number: <u>22496</u>	Calibration Date: <u>3/6/2025</u>
Serial Number: <u>092007</u>	Calibration Due Date: <u>3/6/2026</u>
Model Number: <u>2050W</u>	Manufacturer: <u>Caron</u>

Thermocouple Calibration Check:			°F / °C	°C	
Target Temperature	Unit Reading	Master Thermometer	Accepted Deviation	Actual Deviation	Pass / Fail
20	20	20.8	2.2	0.8	Pass
10	9.8	10.4	2.2	0.6	Pass
0	0	0.1	2.2	0.1	Pass

Master Thermocouple Information: 709

Model Number: <u>4000-90080-09</u>	Calibration Due Date: <u>1/9/2026</u>
Serial Number: <u>240024945</u>	

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay Date: 3/6/2025

Approved By: Samantha Korenek Date: 3/24/2025