



Order Number:	22548
Serial # / Local I.D. #:	5309770000001
Calibration Date:	3/26/2025
Recommended Due Date:	3/26/2026
Certificate Number:	530977000000145742
Condition Found:	Out of Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

120-75 Atmospheric Consistometer 115V Calibration Report

Thermocouple Calibration Check			°F or °C		°C	
Dry Block Temperature	26.7	60.0	87.8	Allowed Deviation	Actual Deviation	Pass / Fail
Controller	27	60	88	± 1.0	0.3	Pass
	Pass	Pass	Pass			
Dry Block I.D. Number	674		Certificate Number	C-43025		
Model Number	ETC-400A 11		Calibration Date	10/23/2024		
Serial Number	692360-02363		Calibration Due Date	10/23/2025		

Motor Speed Accuracy Check					
Target Speed	Measured Tachometer Speed		Allowed Deviation	Actual Deviation	Pass / Fail
150 RPM	152 RPM		± 15 RPM	2 RPM	Pass
	Pass				
Tachometer I.D. #	710		Certificate Number	5-HZ7EP-120-1	
Model Number	461920		Calibration Date	5/8/2024	
Serial Number	131105873		Calibration Due Date	5/8/2025	

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Calibrated By: Amir Mehtal Date: 3/26/2025

Approved By: Samantha Korenek Date: 3/26/2025