



Order Number:	22588
Serial # / Local I.D. #:	522010000001
Calibration Date:	4/15/2025
Recommended Due Date:	4/15/2026
Certificate Number:	52201000000145762
Condition Found:	Out of Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

294-50 Capillary Suction Timer Calibration Report

Balance Calibration Check **					
Target Weight	Reading (Seconds)	Temperature (F°)	Allowed Deviation	Actual Deviation	Pass / Fail
Reading 1	12.9 Sec	73.3 F°	± 3.6 Sec	0.3 Sec	Pass
	Pass	Pass			
Reading 2	12.7 Sec	73.3 F°	± 3.6 Sec		
	Pass	Pass			
Reading 3	13.2 Sec	73.3 F°	± 3.6 Sec		
	Pass	Pass			
Calibration Fluid Information *		Calibration Fluid Lot #	Calibration Fluid Exp.Date		
10000 µS		4061201	6/30/2025		
Thermometer I.D.		709	Certificate Number	4000-14742432	
Model Number	4000-90080-09		Calibration Date	1/9/2024	
Serial Number	240024945		Calibration Due Date	1/9/2026	

* Total dissolved of solids = 6600 Parts Per Million (PPM)

** Three tests of the same temperature should produce a deviation of no more than ± 3.6 seconds.

Calibrated By: Magdy Beshay Date: 4/15/2025

Approved By: Samantha Korenek Date: 4/23/2025