



Roller Oven Calibration Report

Order Number:	538763
Serial # / Local I.D. #:	08-02
Calibration Date:	4/15/2025
Recommended Due Date:	4/15/2026
Certificate Number:	08-0245762
Manufacturer:	OFI Testing Equipment
Part Number:	173-00-C

Condition Found:	In Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	No
(F) or (C):	(°F)
Max Temp:	400
As Found Required:	No

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Calibration Equipment						
Thermometer	336	EA10	171103418	C-40409	8/22/2024	8/22/2025
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Tachometer	312	461920	12050313	C-40454	8/20/2024	8/20/2025
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	345	RH300	Z395285	C-40456	8/20/2024	8/20/2025
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		70	(°F)	Lab / Room Humidity	
						43.0 %

AS LEFT MEASUREMENTS		(°F)	Max Oven Temp	400	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
100.0	104.0	5	4.0	Pass	
	Pass				
200.0	200.0	5	0.0	Pass	
	Pass				
300.0	302.0	5	2.0	Pass	
	Pass				
400.0	403.0	5	3.0	Pass	
	Pass				
25 RPM	23.8	2	-1.2	Pass	
	Pass				

Calibration Technician: Shawn Maglitto

Date: 4/17/2025

Reviewed By: Samantha Korenek

Review Date: 4/23/2025

