



ISCO Pump Verification Report

Order Number: 22624
 Serial Number: 210D20163
 Model: 260D HP

Verification Date: 5/1/2025
 Reverification Date: 5/1/2026

Pressure Check					
Target Pressure	Unit Gauge	Master Gauge Reading	Accepted Tolerance	Actual Deviation	Pass / Fail
1000 PSI	1000.00 PSI	1005.00 PSI	37.50 PSI	5.00 PSI	Pass

Master Gauge Calibration Information: 157

Model Number: JHW60KGZ
 Serial Number: 1391934

Calibration Due Date: 6/19/2025

Leak Check						
Start Pressure	First Reading*	Second Reading**	Duration	Accepted Deviation	Actual Deviation	Pass / Fail
9500	153.640 ml	153.630 ml	30 Minutes	0.01500 ml	0.01000 ml	Pass
* - First reading will be taken at 15 minutes.						
** - Second reading will be taken at 45 minutes. 30 minutes after the 15 minute reading.						

Flow Rate Check		1ml of water is equal to 1g.				
Set Flow Rate (Per Minute)	Initial Volume	Final Volume	Volume Pumped	Volume Collected	Volume Deviation*	Pass / Fail
25.00 g	140.28 g	115.07 g	25.21 g	24.90 g	0.31 g	Pass
* - Volume Deviation can not exceed .25 g / Volume Deviation is Calculated from Volume Pumped and Volume Collected.						

Technician Tested: Magdy Beshay

Date: 5/1/2025

Approved By: Samantha Korenek

Date: 5/13/2025