



131-50 ES Meter Calibration Report

The Device Under Test is certified to meet the accuracy requirements stated in "API 13B-2, 11 Electrical Stability Test".

Equipment Information

Work Order: NEW
Serial Number: 11756

Calibration Date: 5/22/2025
Recommended Calibration Date: 5/22/2026

Calibration Standards

Range: 609V - 1952V

Acceptable tolerance is $\pm 2.5\%$ of full scale (FS) at 25% - 75% FS Range, unless otherwise specified.

These devices have been calibrated using instruments and known standards which are traceable to NIST.

Low Resistor Standard	
Lot Number:	TF656R
Manufacturer:	Caddock®
Description:	HI-Volt Resistor, 9.99 M Ω \pm 0.1%
Reading (V):	609V
Tolerance:	2% (12.18V)

High Resistor Standard	
Lot Number:	TF050R
Manufacturer:	Caddock®
Description:	HI-Volt Resistor, 32.00 M Ω \pm 0.1%
Reading (V):	1952V
Tolerance:	2% (39.04V)

As Left Measurement

	Target	Cycle 1	Variance	Cycle 2	Variance	Cycle 3	Variance	Tolerance	Actual \pm	% Dif	Pass/Fail
Low Resistor	609V	608V	-1V	608V	-1V	608V	-1V	\pm 12.18V	1V	0%	Pass
High Resistor	1952V	1942V	-10V	1942V	-10V	1942V	-10V	\pm 39.04V	10V	1%	Pass
Probe in Water	<3V	2V	N/A	2V	N/A	2V	N/A	N/A	2V	N/A	Pass
Probe in Air	∞	∞	N/A	∞	N/A	∞	N/A	n/a	∞	∞	Pass
Probe Gap (in)	.061in	.061in	.000in	.061in	.000in	.061in	.000in	.001in	.000in	N/A	Pass

Technician: Emad Hanna

Date: 5/22/2025

Approved By: Samantha Korenek

Date: 5/23/2025

Calibrated Per: DCF-LAB-CAL-009