



**131-50 ES Meter Calibration Report**

The Device Under Test is certified to meet the accuracy requirements stated in "API 13B-2, 11 Electrical Stability Test".

**Equipment Information**

<b>Sales Order:</b>	22961	<b>Calibration Date:</b>	5/22/2025
<b>Serial Number:</b>	2812	<b>Recommended Calibration Date:</b>	5/22/2026
<b>Item I.D. #</b>		<b>As Found Data Required by Customer?</b>	<input type="checkbox"/> No <input checked="" type="checkbox"/> Yes
		<b>Repairs Required?</b>	<input type="checkbox"/> No <input checked="" type="checkbox"/> Yes

**Calibration Standards**

Range: 609V - 1952V

Acceptable tolerance is  $\pm 2.5\%$  of full scale (FS) at 25% - 75% FS Range, unless otherwise specified.

These devices have been calibrated using instruments and known standards which are traceable to NIST.

Low Resistor Standard	
<b>Lot Number:</b>	TF050R
<b>Manufacturer:</b>	Caddock®
<b>Description:</b>	HI-Volt Resistor, 9.99 M $\Omega$ $\pm$ 0.1%
<b>Reading (V):</b>	609V
<b>Tolerance:</b>	2% (12.18V)

High Resistor Standard	
<b>Lot Number:</b>	TF656R
<b>Manufacturer:</b>	Caddock®
<b>Description:</b>	HI-Volt Resistor, 32.00 M $\Omega$ $\pm$ 0.1%
<b>Reading (V):</b>	1952V
<b>Tolerance:</b>	2% (39.04V)

**As Left Measurement**

	Target	Cycle 1	Variance	Cycle 2	Variance	Cycle 3	Variance	Tolerance	Actual $\pm$	% Dif	Pass/Fail
<b>Low Resistor</b>	609V	612V	3V	612V	3V	612V	3V	$\pm$ 12.18V	3V	0%	Pass
<b>High Resistor</b>	1952V	1918V	-34V	1918V	-34V	1918V	-34V	$\pm$ 39.04V	34V	2%	Pass
<b>Probe in Water</b>	<3V	2V	N/A	2V	N/A	2V	N/A	N/A	2V	N/A	Pass
<b>Probe in Air</b>	$\infty$	$\infty$	N/A	$\infty$	N/A	$\infty$	N/A	n/a	$\infty$	$\infty$	Pass
<b>Probe Gap (in)</b>	.061in	.061in	.000in	.061in	.000in	.061in	.000in	.001in	.000in	N/A	Pass

**Technician:** Magdy Beshay

**Date:** 5/22/2025

**Approved By:** Samantha Korenek

**Date:** 5/27/2025

Calibrated Per: DCF-LAB-CAL-009