



Order Number:	22759
Serial # / Local I.D. #:	16-522
Calibration Date:	6/24/2025
Recommended Due Date:	6/24/2026
Certificate Number:	16-52245832
Condition Found:	Out of Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

## 100-70 Pressurized Mud Balance Calibration Report

Density Calibration Check				
Target	Reading	Allowed Deviation	Actual Deviation	Pass / Fail
1.00 g/cm <sup>3</sup>	1.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	0.00 g/cm <sup>3</sup>	Pass
	Pass			
2.00 g/cm <sup>3</sup>	2.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	0.00 g/cm <sup>3</sup>	Pass
	Pass			
Ambient Monitor I.D. #	261	Certificate Number	C-47065	
Model Number	EA20	Calibration Date	1/31/2025	
Serial Number	Z375174	Calibration Due Date	1/31/2026	

Ambient Conditions			
Lab Temperature	73.9 F°	Lab Humidity	51.2 %
There is no specific pass or fail criteria per API RP 10B-2. This table is to provide a record of the current ambient conditions during the calibration process.			

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Calibrated By: Magdy Beshay Date: 6/24/2025

Approved By: Samantha Korenek Date: 6/25/2025