



171-49 Temperature Controller Calibration Report

Order Number: 172622 Calibration Date: 6/12/2025
 Serial Number: 526430000001 Recommended Calibration Date: 6/12/2026

| Thermocouple Calibration | | | | | | |
|--------------------------|--------------|---------------|-------------------|---------------|-------------------|-------------|
| Channel | Target Temp. | As Found Temp | Deviation Allowed | Adjusted Temp | Deviation Allowed | Pass / Fail |
| Channel 1 | 120°F | 120°F | ± 4°F | 120°F | ± 0°F | Pass |
| | 450°F | 452°F | ± 10°F | 450°F | ± 0°F | Pass |

Dry Block Calibration Information: 700

Model Number: ETC-400A Calibration Due Date: 3/21/2026
 Serial Number: 704749-02572

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay Date: 6/12/2025

Approved By: Samantha Korenek Date: 07/2/2025