



Order Number:	New
Serial # / Local I.D. #:	5394880000003
Calibration Date:	4/19/2025
Recommended Due Date:	4/19/2026
Certificate Number:	539488000000345766
Condition Found:	NEW
Condition Left:	In Calibration
Blender S/N:	2502778

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

## 120-64-1 Constant Speed Blender 230v Calibration Report

Rotational Speed Test					
Target Speed	Low Point Actual Speed	High Point Actual Speed	Accepted Deviation	Actual Deviation	Pass / Fail
4000 RPM	3995 RPM	4001 RPM	± 100 RPM	6 RPM	Pass
	Pass	Pass			
12000 RPM	11992 RPM	12004 RPM	± 250 RPM	12 RPM	Pass
	Pass	Pass			
Tachometer I.D. Number		680	Certificate Number		C-52606
Model Number		R7050	Calibration Date		7/23/2025
Serial Number		220326689	Calibration Due Date		7/23/2026

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Calibrated By: Elvin Le Date: 8/19/2025

Approved By: Samantha Korenek Date: 8/19/2025