



171-49 Temperature Controller Calibration Report

Order Number: 172622 Calibration Date: 9/19/2025
 Serial Number: 5354590000002 Recommended Calibration Date: 9/19/2026

Thermocouple Calibration						
Channel	Target Temp.	As Found Temp	Deviation Allowed	Adjusted Temp	Deviation Allowed	Pass / Fail
Channel 1	120°F	119°F	± 4°F	120°F	± 0°F	Pass
	450°F	446°F	± 10°F	450°F	± 0°F	Pass

Dry Block Calibration Information: 044

Model Number: ETC400A Calibration Due Date: 6/6/2026
 Serial Number: 623915-01434

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay Date: 9/19/2025

Approved By: Samantha Korenek Date: 9/24/2025