



## Roller Oven Calibration Report

Order Number:	New
Serial # / Local I.D. #:	5397370000001
Calibration Date:	9/26/2025
Recommended Due Date:	9/26/2026
Certificate Number:	539737000000145926
Manufacturer:	OFI Testing Equipment
Part Number:	174-00

Condition Found:	NEW
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	No
( F ) or ( C ):	(°F)
Max Temp:	400
As Found Required:	No

**OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com**

Calibration Equipment							
Thermometer	326	1235C90	181819832	C-46720	2/6/2025	2/6/2026	
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date	
Tachometer	155	4060	122574032	C-52605	7/23/2025	7/23/2026	
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date	
Ambient Monitor	169	EA20	Z375173	C-50047	5/9/2025	5/9/2026	
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date	
Ambient Conditions	Lab / Room Temperature		73.5	(°F)	Lab / Room Humidity		47.6 %

AS LEFT MEASUREMENTS		(°F)	Max Oven Temp	400	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
100.0	102.0	5.0	2.0	Pass	
	Pass				
200.0	198.0	5.0	-2.0	Pass	
	Pass				
300.0	303.0	5.0	3.0	Pass	
	Pass				
400.0	405.0	5.0	5.0	Pass	
	Pass				
25 RPM	25.3	2	0.3	Pass	
	Pass				

Calibration Technician: Emad Hanna

Date: 9/26/2025

Reviewed By: Samantha Korenek

Review Date: 9/26/2025

