



| | |
|--------------------------|--------------------|
| Order Number: | 22784 |
| Serial # / Local I.D. #: | 5167020000001 |
| Calibration Date: | 7/10/2025 |
| Recommended Due Date: | 7/10/2026 |
| Certificate Number: | 516702000000145848 |
| Condition Found: | N/A |
| Condition Left: | In Calibration |
| Part Number: | 112-00-C |
| I.D. Number: | NA |
| Manufacturer: | OFITE |

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

EP and Lubricity Tester Calibration Report

| Tachometer Speed Accuracy Check | | | | |
|---------------------------------|--------------------------|-------------------|----------------------|-------------|
| Target Speed | Tachometer Reading (RPM) | Allowed Deviation | Actual Deviation | Pass / Fail |
| 60.0 RPM | 60.1 RPM | ± 1.5 RPM | 0.1 RPM | Pass |
| | Pass | | | |
| 200.0 RPM | 200.3 RPM | ± 5.0 RPM | 0.3 RPM | Pass |
| | Pass | | | |
| 600.0 RPM | 600.4 RPM | ± 15.0 RPM | 0.4 RPM | Pass |
| | Pass | | | |
| 1000.0 RPM | 1000.3 RPM | ± 25.0 RPM | 0.3 RPM | Pass |
| | Pass | | | |
| Tachometer I.D. # | | 155 | Certificate Number | C-52605 |
| Model Number | | 4060 | Calibration Date | 7/23/2025 |
| Serial Number | | 122574032 | Calibration Due Date | 7/23/2026 |

| Deionized Water Check | | | | |
|-----------------------|----------------------|-------------------|------------------|-------------|
| Target Torque (in-lb) | Unit Reading (in-lb) | Allowed Deviation | Actual Deviation | Pass / Fail |
| 34.0 in-lb | 34.0 in-lb | ± 2.0 in-lb | 0.0 in-lb | Pass |
| | Pass | | | |

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Calibrated By: Emad Hanna Date: 7/10/2025

Approved By: Samantha Korenek Date: 7/10/2025