



## Roller Oven Calibration Report

Order Number:	23009
Serial # / Local I.D. #:	87-30
Calibration Date:	11/18/2025
Recommended Due Date:	11/18/2026
Certificate Number:	87-3045979
Manufacturer:	OFI Testing Equipment
Part Number:	173-00-C

Condition Found:	In Calibration
Condition Left:	In Calibration
Repairs Required:	No
Static Oven:	No
( F ) or ( C ):	(°F)
Max Temp:	400
As Found Required:	No

**OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com**

Calibration Equipment						
Thermometer	154	EA10	Z375331	C-52617	7/23/2025	7/23/2026
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Tachometer	173	461920	131105905	C-46744	2/6/2025	2/6/2026
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Monitor	728	RH300	Z430119	305923	4/1/2025	4/1/2026
	I.D. Number	Model Number	Serial Number	Certificate Number	Calibration Date	Calibration Due Date
Ambient Conditions	Lab / Room Temperature		74.4	(°F)	Lab / Room Humidity	
						48.5 %

AS LEFT MEASUREMENTS		(°F)	Max Oven Temp	400	
Set point	Process Temperature	Accepted Deviation	Actual Deviation	Pass/Fail	
100.0	98.0	5.0	-2.0	Pass	
	Pass				
200.0	199.0	5.0	-1.0	Pass	
	Pass				
300.0	302.0	5.0	2.0	Pass	
	Pass				
400.0	400.0	5.0	0.0	Pass	
	Pass				
25 RPM	25.8	2	0.8	Pass	
	Pass				

Calibration Technician: Shawn Maglitto

Date: 11/18/2028

Reviewed By: Samantha Korenek

Review Date: 11/19/2025

