



Order Number:	23049
Serial # / Local I.D. #:	45928
Calibration Date:	12/18/2025
Recommended Due Date:	12/18/2026
Certificate Number:	4592846009
Condition Found:	N/A
Condition Left:	In Calibration
Part Number:	111-00
I.D. Number:	NA
Manufacturer:	OFITE



11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

EP and Lubricity Tester Calibration Report

Tachometer Speed Accuracy Check				
Target Speed	Tachometer Reading (RPM)	Allowed Deviation	Actual Deviation	Pass / Fail
60.0 RPM	59.5 RPM	± 1.5 RPM	0.5 RPM	Pass
	Pass			
200.0 RPM	199.0 RPM	± 5.0 RPM	1.0 RPM	Pass
	Pass			
600.0 RPM	596.0 RPM	± 15.0 RPM	4.0 RPM	Pass
	Pass			
1000.0 RPM	992.0 RPM	± 25.0 RPM	8.0 RPM	Pass
	Pass			
Tachometer I.D. #		155	Certificate Number	C-52605
Model Number		4060	Calibration Date	7/23/2025
Serial Number		122574032	Calibration Due Date	7/23/2026

Deionized Water Check				
Target Torque (in-lb)	Unit Reading (in-lb)	Allowed Deviation	Actual Deviation	Pass / Fail
34.0 in-lb	34.3 in-lb	± 2.0 in-lb	0.3 in-lb	Pass
	Pass			

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

<u>Emad Hanna</u> Calibration Technician	 <hr/> Technician Signature	<u>12/18/2025</u> Date
<u>Samantha Korenek</u> Reviewed By	 <hr/> Reviewer Signature	<u>12/30/2025</u> Date