



Dependable Products From People You Trust

OFI TESTING EQUIPMENT, INC.

11302 Steeplecrest Drive, Houston, Texas 77065 Phone (832) 320-7300 Fax (713) 880-9886 Web www.ofite.com


CALIBRATION CERTIFICATE

Temperatures are calibrated to meet or exceed the requirements stated in API RP 13B-1 and API RP 13B-2

AS FOUND CHART

AS LEFT CHART

LAB TEMPERATURE	LAB HUMIDITY	PART NUMBER	SERIAL NUMBER	SET POINT	RESULT	ALLOWABLE RANGE	CALIBRATION DATE	DUE DATE	TECHNICIAN
70.00 °F	23.00 %	165-14-1	5402370000003	930.0 °F	977.6 °F	860-1000 °F	12/30/2025	12/30/2027	SM



Calibration station has been calibrated using instruments which are traceable to NIST.

Channel 3

CAL STATION SERIAL NUMBER: 6B-11 **CAL STATION CALIBRATION DATE:** 11/04/2025 **CAL STATION DUE DATE:** 11/04/2026
AMBIENT MONITOR MODEL: RH300 **AMBIENT MONITOR S/N:** Z430119 **AMBIENT MONITOR CAL DATE:** 04/01/2025 **AMBIENT MONITOR DUE DATE:** 04/01/2026

UNCERTAINTY: ± 2.8°C / 5°F **CALIBRATION PROCEDURE USED** Retort Calibration Procedures API RP 13B-1

AS FOUND CONDITION:

CALIBRATION RESULT: Accepted