



OFI Testing Equipment, Inc.
 11302 Steeplecrest Drive Houston, TX 77065 USA
 Phone: 832-320-7300 · Fax: 713-880-9886 · Web: www.ofite.com

Viscometer Calibration Report

Oil Calibration Per API Recommended Practice 13B-1 Annex H.5 & 13B-2 Annex J.5 Procedure - Calibration of Viscometers

Repairs Required? No

As Found Data Required by Customer? No

Part Number: 132-00

Serial Number: 5396230000004

S/O# or I/O# New

Bob B1 Spring F1

All master instruments and fluids are calibrated and/or certified to standards which are traceable to NIST.

Instrument Certificates of Calibration attached

Fluid Thermometer: 704

Model #: 1000,90080-0

Calibration Date: 5/15/2025

Calibration Due: 5/15/2026

Serial No: 221540132

Tachometer: 155

Model #: 4060

Calibration Date: 7/23/2025

Calibration Due: 7/23/2026

Serial No: 122574032

Ambient Monitor: 169

Model #: EA20

Calibration Date: 5/9/2025

Calibration Due: 5/9/2026

Serial No: Z375173

Ambient Conditions

Room Temperature: 72.5

Relative Humidity: 46.9

RPM's:

Speed Setting: 300 RPM Actual: 598.5

Speed Setting: 600 RPM Actual: 598

AS LEFT MEASUREMENTS

Calibration Fluid Viscosity	Calibration Fluid Batch #	Temp	Calibrated Viscosity (CV)	600 RPM Reading	600 VIS: 600 Reading * 0.5050	Accepted Deviation	Actual Deviation (600 VIS - CV)	300 RPM Reading	300 VIS: 300 Reading * 1.0000	Accepted Deviation	Actual Deviation	Pass / Fail
100 cP	508512	21.4	103.8	208	105.05	±1.50	1.25	103.5	103.50	±1.50	-0.30	Pass
50 cP	228501	20.8	52.3	103.5	52.27	±1.50	-0.03	51.5	51.50	±1.50	-0.80	Pass
											Calibration Result	Pass

Technician: Emad Hanna

Uncertainty: 0.5%

Reviewed By: Samantha Korenek

Calibration Date: 1/6/2026

Calibration Due : 1/6/2027

Review Date: 1/6/2026

Calibration Fluid Activation Date: 100 cP 1/6/2026 50 cP 1/6/2026

(Activation Date should not exceed 30 days of the calibration date)