




Order Number:	176270
Serial # / Local I.D. #:	5264300000002
Calibration Date:	1/5/2025
Recommended Due Date:	1/5/2026
Certificate Number:	526430000000245662
Condition Found:	
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

171-49 Temperature Controller Calibration Report

Thermocouple Verification Check						
Channel	Target Temperature	As Found Temperature	Deviation Allowed	Adjusted Temperature	Deviation Allowed	Pass / Fail
Channel 1	120.0 °F	120.0 °F	± 4.0 °F	120.0 °F	± 0.0 °F	Pass
	450.0 °F	456.0 °F	± 10.0 °F	450.0 °F	± 0.0 °F	Pass
Channel 1 Thermocouple Serial Number or I.D. Number*				NA		
Dry Block I.D. Number		131	Certificate Number		C-47066	
Model Number		ETC400A	Calibration Date		2/12/2025	
Serial Number		604975-01189	Calibration Due Date		2/12/2026	
*If Applicable - If the Thermocouple is not serialized then enter N/A in the box .						

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

<u>Magdy Beshay</u> Calibration Technician	 <hr/> Technician Signature	<u>1/5/2025</u> Date
<u>Samantha Korenek</u> Reviewed By	 <hr/> Reviewer Signature	<u>1/6/2026</u> Date