



Order Number:	23098
Serial # / Local I.D. #:	GY-P004
Calibration Date:	1/21/2026
Recommended Due Date:	1/21/2027
Certificate Number:	GY-P00446043
Condition Found:	Out of Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

## 100-70 Pressurized Mud Balance Calibration Report

Density Calibration Check				
Target	Reading	Allowed Deviation	Actual Deviation	Pass / Fail
1.00 g/cm <sup>3</sup>	1.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	0.00 g/cm <sup>3</sup>	Pass
	Pass			
2.00 g/cm <sup>3</sup>	2.00 g/cm <sup>3</sup>	± 0.01 g/cm <sup>3</sup>	0.00 g/cm <sup>3</sup>	Pass
	Pass			
Ambient Monitor I.D. #	728	Certificate Number	305923	
Model Number	RH300	Calibration Date	4/1/2025	
Serial Number	Z430119	Calibration Due Date	4/1/2026	

Ambient Conditions			
Lab Temperature	69.2 F°	Lab Humidity	45.6 %
There is no specific pass or fail criteria per API RP 10B-2. This table is to provide a record of the current ambient conditions during the calibration process.			

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Magdy Beshay	<i>Magdy Beshay</i>	1/21/2026
Calibration Technician	Technician Signature	Date
Samantha Korenek	<i>Samantha Korenek</i>	1/22/2026
Reviewed By	Reviewer Signature	Date