



Order Number:	23173
Serial # / Local I.D. #:	5234340000016
Verification Date:	3/5/2026
Recommended Due Date:	3/5/2027
Certificate Number:	523434000001646086
Condition Found:	N/A
Condition Left:	In Tolerance
As Found Required:	No

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Pressure Gauge Verification Report

As Left		Gauge Location / Assembly Part Number		170-04	
Max Pressure on Unit Gauge (PSI)				1000 PSI	
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
250 PSI	253 PSI	250 PSI	± 10 PSI	-3 PSI	Pass
500 PSI	504 PSI	500 PSI	± 10 PSI	-4 PSI	Pass
750 PSI	754 PSI	750 PSI	± 10 PSI	-4 PSI	Pass
Master Gauge I.D. Number		129	Certificate Number		C-55944
Model Number		DPG1000B2000PSIG-5	Calibration Date		10/16/2025
Serial Number		5186401001	Calibration Due Date		10/16/2026

The gauge tested above was tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, this gauge is verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

<u>Magdy Beshay</u> Verification Technician	<u>3/5/2026</u> Date
<u>Samantha Korenek</u> Reviewed By	<u>3/10/2026</u> Date