



Order Number:	23015
Serial # / Local I.D. #:	5250100000001
Calibration Date:	2/25/2026
Recommended Due Date:	2/25/2027
Certificate Number:	525010000000146078
Condition Found:	Out of Calibration
Condition Left:	In Calibration
Cell Serial #:	5250100000001

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

## 120-70 Stirred Fluid Loss Tester - 115V Calibration Report

Thermocouple Calibration Check			°F or °C		°F	
Dry Block Temperature	85	200.0	395.0	Allowed Deviation	Actual Deviation	Pass / Fail
Cell	86	201	396	2	1.0	Pass
	Pass	Pass	Pass			
Dry Block I.D. Number		44	Certificate Number		C-51136	
Model Number	ETC400A		Calibration Date		6/6/2025	
Serial Number	623915-01434		Calibration Due Date		6/6/2026	

Tachometer Speed Accuracy Check						
Unit Display Reading	150 RPM			Allowed Deviation	Actual Deviation	Pass / Fail
Tachometer Reading	151 RPM			± 15 RPM	1 RPM	Pass
	Pass					
Tachometer I.D. Number		680	Certificate Number		C-52606	
Model Number	R7050		Calibration Date		7/23/2025	
Serial Number	220326689		Calibration Due Date		7/23/2026	

Low Pressure Gauge Calibration Check					
Target Pressure	Master Gauge Reading	Unit Reading	Allowed Deviation	Actual Deviation	Pass / Fail
250	248 PSI	250 PSI	± 50 PSI	2 PSI	Pass
	Pass	Pass			
500	495 PSI	500 PSI	± 50 PSI	5 PSI	Pass
	Pass	Pass			
750	752 PSI	750 PSI	± 50 PSI	2 PSI	Pass
	Pass	Pass			
Master Gauge I.D. Number		143	Certificate Number		C-58776
Model Number	M1 series		Calibration Date		1/6/2026
Serial Number	251359		Calibration Due Date		1/6/2027

High Pressure Gauge Calibration Check					
Target Pressure	Master Gauge Reading	Unit Reading	Allowed Deviation	Actual Deviation	Pass / Fail
500	512 PSI	500 PSI	± 50 PSI	12 PSI	Pass
	Pass	Pass			
1000	1006 PSI	1000 PSI	± 50 PSI	6 PSI	Pass
	Pass	Pass			
2000	1995 PSI	2000 PSI	± 50 PSI	5 PSI	Pass
	Pass	Pass			
Master Gauge I.D. Number		143	Certificate Number		C-58776
Model Number		M1 series	Calibration Date		1/6/2026
Serial Number		251359	Calibration Due Date		1/6/2027

Timer Accuracy Check					
Target Time in Minutes (hh:mm:ss)	Calibrated Timer (hh:mm:ss)	Unit Timer (hh:mm:ss)	Allowed Deviation	Actual Deviation	Pass / Fail
00:12:00	00:12:00	00:12:00	± 1 Sec	0 Sec	Pass
		Pass			
Calibrated Timer I.D. Number		713	Certificate Number		5013-15156497
Model Number		5013,94461-15	Calibration Date		5/30/2024
Serial Number		240436643	Calibration Due Date		5/30/2026

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Magdy Beshay  
Calibration Technician

*Magdy Beshay*  
Technician Signature

2/25/2026  
Date

Samantha Korenek  
Reviewed By

*Samantha Korenek*  
Reviewer Signature

3/11/2026  
Date