



Order Number:	23225
Serial # / Local I.D. #:	5280850000001
Verification Date:	3/31/2026
Recommended Due Date:	3/31/2027
Certificate Number:	528085000000146112
Condition Found:	N/A
Condition Left:	In Tolerance
As Found Required:	No

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

## Pressure Gauge Verification Report

<b>As Left</b>		Gauge Location / Assembly Part Number		170-06	
Max Pressure on Unit Gauge (PSI)				200 PSI	
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
50 PSI	50 PSI	50 PSI	± 2 PSI	0 PSI	Pass
100 PSI	101 PSI	100 PSI	± 2 PSI	-1 PSI	Pass
150 PSI	151 PSI	150 PSI	± 2 PSI	-1 PSI	Pass
Master Gauge I.D. Number		166	Certificate Number		C-58778
Model Number		68935-22	Calibration Date		2026-01-06
Serial Number		5186402001	Calibration Due Date		2027-01-06

The gauge tested above was tested for compliance with API 13A, 5.2.7 "Laboratory pressure-measuring device". When passed, this gauge is verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

<u>Magdy Beshay</u> Verification Technician	<u>3/31/2026</u> Date
<u>Samantha Korenek</u> Reviewed By	<u>3/31/2026</u> Date