



Order Number:	23230
Serial # / Local I.D. #:	GY-P001
Calibration Date:	4/8/2026
Recommended Due Date:	4/8/2027
Certificate Number:	GY-P00146120
Condition Found:	In Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

## 100-70 Pressurized Mud Balance Calibration Report

Density Calibration Check				
Target	Reading	Allowed Deviation	Actual Deviation	Pass / Fail
1.00 g/cm3	1.00 g/cm3	± 0.01 g/cm3	0.00 g/cm3	Pass
	Pass			
2.00 g/cm3	2.00 g/cm3	± 0.01 g/cm3	0.00 g/cm3	Pass
	Pass			
Ambient Monitor I.D. #	169	Certificate Number	C-50047	
Model Number	EA20	Calibration Date	5/9/2025	
Serial Number	Z375173	Calibration Due Date	5/9/2026	

Ambient Conditions			
Lab Temperature	73.9 F°	Lab Humidity	43.8 %
There is no specific pass or fail criteria per API RP 10B-2. This table is to provide a record of the current ambient conditions during the calibration process.			

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

B.T.M _____ Calibration Technician	 _____ Technician Signature	4/8/2026 _____ Date
Samantha Korenek _____ Reviewed By	 _____ Reviewer Signature	4/9/2026 _____ Date