



Chiller Calibration Report

Order Number: <u>0023231</u>	Calibration Date: <u>4/8/2026</u>
Serial Number: <u>2050-1-1-153</u>	Calibration Due Date: <u>4/8/2027</u>
Model Number: <u>152-55</u>	Manufacturer: <u>Caron 2050</u>

Thermocouple Calibration Check:			°F / °C	°C	
Target Temperature	Unit Reading	Master Thermometer	Accepted Deviation	Actual Deviation	Pass / Fail
20	20.2	19	2.2	1.2	Pass
10	10.2	10	2.2	0.2	Pass
0	-0.4	-0.2	2.2	0.4	Pass

Master Thermocouple Information: 260

Model Number: <u>90080-09</u>	Calibration Due Date: <u>1/6/2027</u>
Serial Number: <u>192283369</u>	

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Terrill Braun Date: 4/8/2026

Approved By: Samantha Korenek Date: 4/9/2026