



Order Number:	23281
Serial # / Local I.D. #:	13-971
Calibration Date:	4/22/2026
Recommended Due Date:	4/22/2027
Certificate Number:	13-97146134
Condition Found:	In Calibration
Condition Left:	In Calibration

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

100-70 Pressurized Mud Balance Calibration Report

Density Calibration Check				
Target	Reading	Allowed Deviation	Actual Deviation	Pass / Fail
1.00 g/cm ³	1.00 g/cm ³	± 0.01 g/cm ³	0.00 g/cm ³	Pass
	Pass			
2.00 g/cm ³	2.00 g/cm ³	± 0.01 g/cm ³	0.00 g/cm ³	Pass
	Pass			
Ambient Monitor I.D. #	169	Certificate Number	C-50047	
Model Number	EA20	Calibration Date	5/9/2025	
Serial Number	Z375173	Calibration Due Date	5/9/2026	

Ambient Conditions			
Lab Temperature	73.9 F°	Lab Humidity	43.8 %
There is no specific pass or fail criteria per API RP 10B-2. This table is to provide a record of the current ambient conditions during the calibration process.			

This report meets or exceeds the requirements outlined in API RP 10B-2 Recommended Practice for Testing Well Cements for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

<u>Arnulfo Muniz</u> Calibration Technician	 <hr/> Technician Signature	<u>4/22/2026</u> Date
<u>Samantha Korenek</u> Reviewed By	 <hr/> Reviewer Signature	<u>4/22/2026</u> Date