Order Number:	23285
Serial # / Local I.D. #:	T10307
Verification Date:	4/16/2026
Recommended Due Date:	3/31/2027
Certificate Number:	
Condition Found:	Out of Tolerance
Condition Left:	In Tolerance
As Found Required:	No

OFITE®
OFI Testing Equipment, Inc.

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Pressure Gauge
Verification Report

As Left	Gauge Location / Assembly Part Number			170-04	
			Max Pressure on Unit Gauge (PSI)		1000 PSI
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
250 PSI	256 PSI	250 PSI	± 10 PSI	-6 PSI	Pass
500 PSI	507 PSI	500 PSI	± 10 PSI	-7 PSI	Pass
750 PSI	752 PSI	750 PSI	± 10 PSI	-2 PSI	Pass
Master Guage I.D. Number		129	Certificate Number		C-55944
Model Number		DPG1000B2000PSIG-5	Calibration Date		10/16/2025
Serial Number		5186401001	Calibration Due Date		10/16/2026

The gauge tested above was tested for compliance with API 13A, 5.2.7 “Laboratory pressure-measuring device”. When passed, this gauge is verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Arnulfo Muniz	4/16/2026
Verification Technician	Date
Samantha Korenek	4/23/2026
Reviewed By	Date