Order Number:	23286
Serial # / Local I.D. #:	05B03A7M
Verification Date:	4/16/2026
Recommended Due Date:	3/31/2027
Certificate Number:	
Condition Found:	Out of Tolerance
Condition Left:	In Tolerance
As Found Required:	No

11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

Pressure Gauge
Verification Report

As Left	Gauge Location / Assembly Part Number			170-06	
			Max Pressure on Unit Gauge (PSI)		200 PSI
Target Pressure	Master Gauge Reading	Unit Gauge Reading	Allowed Deviation	Actual Deviation	Pass / Fail
50 PSI	48 PSI	50 PSI	± 2 PSI	2 PSI	Pass
100 PSI	101 PSI	100 PSI	± 2 PSI	-1 PSI	Pass
150 PSI	148 PSI	150 PSI	± 2 PSI	2 PSI	Pass
Master Guage I.D. Number		166	Certificate Number		C-58778
Model Number		68935-22	Calibration Date		1/6/2026
Serial Number		5186402001	Calibration Due Date		1/6/2027

The gauge tested above was tested for compliance with API 13A, 5.2.7 “Laboratory pressure-measuring device”. When passed, this gauge is verified to measure within 2 ½% of Full Scale when compared to an NIST traceable gauge, measured at 25%, 50% and 75% of Full Range.

Arnulfo Muniz	4/16/2026
Verification Technician	Date
Samantha Korenek	4/29/2026
Reviewed By	Date