



Roller Oven Calibration Report

| | |
|--------------------------|-----------------------|
| Order Number: | RP0000001 |
| Serial # / Local I.D. #: | 08-67 |
| Calibration Date: | 5/20/2026 |
| Recommended Due Date: | 5/20/2027 |
| Certificate Number: | 08-6746162 |
| Manufacturer: | OFI Testing Equipment |
| Part Number: | 174-00 |

| | |
|--------------------|----------------|
| Condition Found: | In Calibration |
| Condition Left: | In Calibration |
| Repairs Required: | No |
| Static Oven: | No |
| (F) or (C): | (°F) |
| Max Temp: | 400 |
| As Found Required: | No |

OFI Testing Equipment, Inc. - 11302 Steeplecrest Dr - Houston, Texas - 77065 - Phone: 832.320.7300 - Web: www.ofite.com

| Calibration Equipment | | | | | | |
|-----------------------|------------------------|--------------|---------------|--------------------|---------------------|----------------------|
| Thermometer | 154 | EA10 | Z375331 | C-52617 | 7/23/2025 | 7/23/2026 |
| | I.D. Number | Model Number | Serial Number | Certificate Number | Calibration Date | Calibration Due Date |
| Tachometer | 706 | R7050 | 220502483 | C-54329 | 9/5/2025 | 9/5/2026 |
| | I.D. Number | Model Number | Serial Number | Certificate Number | Calibration Date | Calibration Due Date |
| Ambient Monitor | 261 | EA20 | Z375174 | C-59375 | 1/20/2026 | 1/20/2027 |
| | I.D. Number | Model Number | Serial Number | Certificate Number | Calibration Date | Calibration Due Date |
| Ambient Conditions | Lab / Room Temperature | | 70 | (°F) | Lab / Room Humidity | |
| | | | | | | 41.0 % |

| AS LEFT MEASUREMENTS | | (°F) | Max Oven Temp | 400 | |
|----------------------|---------------------|--------------------|------------------|-----------|--|
| Set point | Process Temperature | Accepted Deviation | Actual Deviation | Pass/Fail | |
| 100.0 | 101.0 Pass | 5.0 | 1.0 | Pass | |
| 200.0 | 201.0 Pass | 5.0 | 1.0 | Pass | |
| 300.0 | 300.0 Pass | 5.0 | 0.0 | Pass | |
| 400.0 | 399.0 Pass | 5.0 | -1.0 | Pass | |
| 25 RPM | 25.7 Pass | 2 | 0.7 | Pass | |
| | | | | Pass | |

Calibration Technician: Shawn Maglitto

Technician Signature: *Shawn Maglitto*

Date: 5/20/2026

Reviewed By: _____

Reviewer Signature: Samantha Korenek

Review Date: 5/21/2026

