



Chiller Calibration Report

Order Number: RP00000028
Serial Number: 2050-1-1-141
Model Number: 152-55

Calibration Date: 5/21/2026
Calibration Due Date: 5/21/2027
Manufacturer: CARON

Thermocouple Calibration Check:			°F / °C	°C	
Target Temperature	Unit Reading	Master Thermometer	Accepted Deviation	Actual Deviation	Pass / Fail
20	20.3	20.1	2.2	0.3	Pass
10	10.1	10.6	2.2	0.6	Pass
0	0.5	0.5	2.2	0.5	Pass

Master Thermocouple Information: 325

Model Number: 90080-09
Serial Number: 192283357

Calibration Due Date: 7/23/2026

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST.

Technician Tested: Magdy Beshay

Date: 5/21/2026

Approved By: Samantha Korenek

Date: 5/27/2026