



Silverson Mixer Calibration Report

Order Number: RO00000004
Part Number: 152-15
Serial Number: 41471
I.D. Number:

Verification Date: 5/28/2026
Verification Due Date: 5/28/2027

Rotational Speed Test

Target Speed	Unit Display Speed	Tachometer Speed	Allowed Deviation	Actual Deviation	Pass / Fail
2500 RPM	2500 RPM	2516 RPM	± 300 RPM	16 RPM	Pass
5000 RPM	5000 RPM	5014 RPM	± 300 RPM	14 RPM	Pass
7500 RPM	7500 RPM	7526 RPM	± 300 RPM	26 RPM	Pass
9200 RPM	9200 RPM	9239 RPM	± 300 RPM	39 RPM	Pass

Calibrated Tachometer I.D. 680

Model Number: R7050
Serial Number: 220326689

Calibration Due Date: 7/23/2026

Calibrated By: Magdy Beshay

Date: 5/28/2026

Approved By: Samantha Korenek

Date: 5/28/2026

