



Order Number:	RP00000035
Serial # / Local I.D. #:	46363
Calibration Date:	6/12/2026
Recommended Due Date:	6/12/2027
Certificate Number:	4636346185
Condition Found:	Out of Calibration
Condition Left:	In Calibration
Part Number:	112-00
I.D. Number:	
Manufacturer:	OFITE

11302 Steeplecrest Dr - Houston, Texas - 77064 - Phone: 832.320.7300 - E-Mail: www.ofite.com

EP and Lubricity Tester Calibration Report

Tachometer Speed Accuracy Check				
Target Speed	Tachometer Reading (RPM)	Allowed Deviation	Actual Deviation	Pass / Fail
60.0 RPM	60.5 RPM	± 1.5 RPM	0.5 RPM	Pass
	Pass			
200.0 RPM	200.8 RPM	± 5.0 RPM	0.8 RPM	Pass
	Pass			
600.0 RPM	601.0 RPM	± 15.0 RPM	1.0 RPM	Pass
	Pass			
1000.0 RPM	997.1 RPM	± 25.0 RPM	2.9 RPM	Pass
	Pass			
Tachometer I.D. #		680	Certificate Number	C-52606
Model Number		R7050	Calibration Date	7/23/2025
Serial Number		220326689	Calibration Due Date	7/23/2026

Deionized Water Check				
Target Torque (in-lb)	Unit Reading (in=lb)	Allowed Deviation	Actual Deviation	Pass / Fail
34.0 in-lb	33.4 in-lb	± 2.0 in-lb	0.6 in-lb	Pass
	Pass			

This report meets or exceeds the requirements outlined in API SP 13A Specification for Drilling-Fluid Materials for calibration of laboratory devices, using certified master test equipment that is traceable to NIST which is a signatory to the Mutual Recognition Arrangement (MRA) with the International Laboratory Accreditation Cooperation (ILAC).

Calibrated By: Magdy Beshay Date: 6/12/2026

Approved By: Samantha Korenek Date: 6/15/2026